

## CERTIFICADO N° 025L-CMQ-2021 v.1



El Laboratorio de Investigación y Control de Calidad en Textiles Cuero y Calzado de la Universidad de Santiago LICTEX-UdeSantiago certifica que

**UN TOTAL DE 3.000 cajas**  
**MASCARILLAS desechables**  
**Identificadas como MMK0001 hasta MMK 3000**

Suministrado por  
**Importadora, Exportadora y Distribuidora B y B LTDA.**  
**Avenida Los Carrera 1676, Concepción**

Elaborado en  
**Avenida Los Carrera 1676, Concepción**  
**ES CONFORME CON UN**  
**99% FILTRACIÓN BACTERIANA (BFE)**

NORMAS QUE APLICA:

**Especificación UNE 0064-1:2020 abril**

Para conceder este certificado, LICTEX-UdeSantiago ha ensayado el producto cuyos resultados se evidencian en el Informe de ensayos

**N° 39L-CMQ21 v.1 06-08-2021**

**Fecha de concesión v.1 07-agosto-2021**  
**v.0 06-agosto-2021**

**Dra KARINA BRAVO FLORES**  
BIOQUIMICA  
JEFA LABORATORIO CMQ  
LICTEX – UdeSantiago



**MARIA NEIRA LORCA**  
INGENIERA TEXTIL  
GERENTE TECNICO  
LICTEX – UdeSantiago

---

**CERTIFICADO N° 025L-CMQ-2021 v.1**

---



# Certificate of Compliance



No. 5C200413.YCKTQ99

Certificate's  
Holder:

Yushan County Kust Medical Co., Ltd.  
Yiyang County Jiangxi Yushan County High-Tech  
Zone 10 Lang Line

Certification ECM  
Mark:



Product:  
Model(s):

KN95 Protective Mask  
KN95 FFP2

Verification to:

Standard:  
EN 149:2001+A1:2009

related to CE Directive(s):  
R 2016/425 (Personal Protective Equipment)

**Remark:** This document has been issued on a voluntary basis and upon request of the manufacturer. It is our opinion that the technical documentation received from the manufacturer is satisfactory for the requirements of the ECM Certification Mark. The conformity mark above can be affixed on the products accordingly to the ECM regulation about its release and its use.

Additional information and clarification about the Marking:



The manufacturer is responsible for the CE Marking process, and if necessary, must refer to a Notified Body. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products, RG01\_ECM rev.3 available at: [www.entecerma.it](http://www.entecerma.it)

**Issuance date: 07 April 2020**

**Expiry date: 06 April 2025**

Reviewer  
Technical expert  
Amanda Payne

Approver  
ECM Service Director  
Luca Bedonni

Ente Certificazione Macchine Srl

Via Ca' Bella, 243 - Loc. Castello di Serravalle - 40053 Valsamoggia (BO) - ITALY  
☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it



180011112242



(2018) 国认监认字(244)号



中国认可  
国际互认  
TESTING  
CNAS L6780

# 检验报告

## TEST REPORT



扫一扫关注我们



扫一扫查询真伪



报告编号

REPORT NO.

国纺委字第 W202004745 号

产品名称

NAME OF SAMPLE

KN95 口罩

委托单位

CUSTOMER

玉山县酷斯特医疗科技有限公司

检验类别

TEST CATEGORY

委托检验

浙江省轻工业品质量检验研究院

(浙江省纺织测试研究院)

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心(浙江)

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)



## QUALIFICATION CERTIFICATE

# 合格证

产品名称: KN95口罩(非医用)

主要成份: 70%无纺布, 30%熔喷布

Main ingredients: 70% non-woven, 30% melt

型号:耳戴式 / Model: Earwear type

数量(只):10 / Quantity: 10 pcs

检验员: 08 / Inspector: 08

生产批号: 20200320

Production batch number: 20200320

生产日期: / Date of production:

到期日: / Date of Expiry:

执行标准: GB 2626-2006

Implementation Standard: GB 2626-2006

有效期:二年 / Validity period: two years

存储条件:常温存储/相对湿度 $\leq$ 80%

Storage conditions: storage at room temperature, relative humidity $\leq$ 80%

注意事项:本产品不能用于发热门诊, 隔离病房, 隔离留观病房, 手术室及隔离重症监护室等区域。

Notice: this product can not be used in fever clinic, isolation ward, isolation observation ward, operating room and isolation intensive care unit and other areas.

玉山县酷斯特医疗科技有限公司

YUSHAN KUSTER MEDICAL TECHNOLOGY CO., LTD.

江西省上饶市玉山县高新区十坊线(十字坊线2号)

江西创丰光电发展有限公司9号厂房内

Jiangxi province. In the No. 9 factory building of Jiangxi

Chuangfeng Optoelectronic Development Co., Ltd.,

Shifang Line (No. 2 Cross Square Line), High-tech Zone,

Yushan County, Shangrao City.

中国制造 / MADE IN CHINA

尺寸: 45x65mm

浙江省轻工业产品质量检验研究院  
国家纺织服装产品质量监督检验中心（浙江）

# 检验报告

国纺委字第 W202004745 号

第 1 页 共 3 页

委托单位名称 Name of Customer	玉山县酷斯特医疗科技有限公司		地址 Address	—
生产单位 Manufacturer	—		地址 Address	—
样品信息 Sample information	样品名称 Name of sample: KN95 口罩 样品特性 Characteristics: 白色 商标 Trademark: — 规格/号型 Specification/model: — 等级 Level: KN95 安全技术类别 Category of safety specification: — 样品款号/货号 Art. No.: — -----			
以上为客供信息 (Above-mentioned information by Customer-supplied)				
来样方式 The sent way of sample	快递	样品数量 Sample quantity	1 包	
送检日期 Receiving Date of Sample	2020-04-08	检测类别 Test Category	委托检验	
判定依据 Rating Requirements	GB 2626-2006			
检测结论/Test Summary:  实测结果详见附页。				
 <p>批准日期/ Date of Approval: 2020-04-08</p>				
备注 Remarks	样品未经预处理, 数据供参考。			

签发:  
Approved by

张雪芳

# 检验报告

国纺委字第 W202004745 号

第 2 页 共 3 页

序号	检测项目	检测方法	单位	标准要求 (KN95)	实测值	单项评价	结果备注
1	过滤效率 (盐性介质)	GB 2626-2006 6.3	%	$\geq 95.0$	95.8	符合	—
2	呼吸阻力	吸气阻力	Pa	$\leq 350$	64	符合	—
		呼气阻力	Pa	$\leq 250$	50		





FACE MASK FOLDING PARTICULATE SPECIAL PROTECTIVE

# KN95

All protective parts and components are made of high quality non-woven fabric. The elastic ear loops and adjustable headbands are made of high quality elastic material. The outer and inner layers are made of high quality non-woven fabric.



3D  
FOLD DESIGN



- FOLDED PROTECTIVE
- ALLERGY FREE
- BREATHABLE
- SANITARY DISPOSABLE DESIGN

CONT: 50 PCS

NOSE BRIDGE STRIP  
KN95

### GENRE

Highly effective against particulate matter, pollen, dust, mold, bacteria, and viruses. It is suitable for use in high-risk environments.



### NOTICE FOR USE

1. Before use, please check the expiration date and the storage conditions. 2. Please avoid using the mask in high-humidity environments. 3. Please avoid using the mask in high-temperature environments.

4. Please avoid using the mask in high-pollution environments. 5. Please avoid using the mask in high-risk environments. 6. Please avoid using the mask in high-risk environments.

7. Please avoid using the mask in high-risk environments.

8. Please avoid using the mask in high-risk environments.





Thermal non-woven fabric

Filter sponge

1

2

3

4

5

Double layer 25g meltblown cloth

No-woven fabric

Double layer 25g meltblown cloth





